



# ACCREDITATION CERTIFICATE

Issued under the authority of Bangladesh Accreditation Act, 2006  
by Bangladesh Accreditation Board (BAB), Ministry of Industries to

**Parkview Diagnostic Center, Parkview Hospital Ltd.**  
**94/103, Katalgonj Road, Panchlaish**  
**Chattogram-4000, Bangladesh**

This is to certify that this

**Medical Laboratory**

is accredited in accordance with the international standard

**ISO 15189:2012**

in respect of the associated scope, subject to the terms and conditions governing the relevant conformity assessment body (CAB) accreditation.

Certificate Number : 03.009.25  
Accreditation Date : 30 January 2025  
Date of Issuance : 30 January 2025  
Date of Expiration : 29 January 2028



*30.01.2025*  
Md. Anwarul Alam  
Director General

This certificate must be returned on request; reproduction must follow BAB guidelines. For the specific scopes to which this accreditation applies, please refer to the Directory of CABs at BAB website.

## SCOPE OF ACCREDITATION

(For Medical Laboratory)

**CAB Name & Address:** Parkview Diagnostic Center, Parkview Hospital Ltd.  
94/103, Katalgonj Road, Panchlaish, Chattogram 4000, Bangladesh

**Accreditation Standard:** ISO 15189:2012      **Accreditation Date:** 30 January 2025  
**Certificate Number:** 03.009.25      **Issued on:** 30 January 2025  
**Last Amended on:** -      **Valid until:** 29 January 2028  
**Amendment no:** -

S.N.	Products/ Materials/ Items of test	Type of tests performed	Specifications/ Standard test methods/Techniques used	Range of testing/Limit of detection
<b>Field: Clinical Biochemistry</b>				
1.	Blood (Plasma)	Glucose	Hexokinase-G-6-PDH	0–500 mg/dL
2.	Blood (Serum)	Creatinine	Modified kinetic Jaffe, Alkaline picrate, IDMS traceability	0.15–20 mg/dL
3.	Blood (Serum)	Urea	Urease/GLDH	0–150 mg/dL
4.	Blood (Serum)	Uric Acid	Uricase	0–20 mg/dL
5.	Blood (Serum)	Albumin	BCP dye-binding	0.6 – 8 g/dL
6.	Blood (Serum)	Total Bilirubin	Modified Diazo, Jendrassik-Grof	0.1–25 mg/dL
7.	Blood (Serum)	ALT (SGPT)	IFCC, L-Alanine/α-KG/ NADH to NAD/ Calibrated/37°C, P5P	6–1000 U/L
8.	Blood (Serum)	AST (SGOT)	IFCC, L-Aspartate/ α KG/NADH to NAD/ Calibrated/37°C, P5P	0–1000 U/L
9.	Blood (Serum)	Alkaline Phosphatase (ALP)	IFCC, p-NPP/AMP/ Calibrated/37°C	10–1000 U/L
10.	Blood (Serum)	Total Cholesterol	Enzymatic, CE/CO/HPO	50–600 mg/dL
11.	Blood (Serum)	HDL Cholesterol	Liquid selective detergent, PEG-modified CE/CO, PO	3–150 mg/dL
12.	Blood (Serum)	LDL Cholesterol	Liquid selective detergent, CE/CO/PO	5–300 mg/dL
13.	Blood (Serum)	Triglyceride	Enzymatic-colorimetric, LPL/GK/GPO/POD	15–1000 mg/dL
14.	Blood (Plasma /Serum)	Sodium	Indirect ISE	50–200 mmol/L
15.	Blood (Plasma /Serum)	Potassium	Indirect ISE	1–10 mmol/L
16.	Blood (Plasma /Serum)	Chloride	Indirect ISE	50–200 mmol/L
17.	Blood (Serum)	Calcium	o-Cresolphthalein complexone (oCPC)	5.0–15 mg/dL
18.	Blood (Serum)	TSH	CLIA	0.010–150 mIU/L
19.	Blood (Serum)	Free T4	CLIA	1.3–155 pmol/L

  
Quality Manager

## SCOPE OF ACCREDITATION

(For Medical Laboratory)

**CAB Name & Address:** Parkview Diagnostic Center, Parkview Hospital Ltd.  
 94/103, Katalgonj Road, Panchlaish, Chattogram 4000, Bangladesh

**Accreditation Standard:** ISO 15189:2012      **Accreditation Date:** 30 January 2025  
**Certificate Number:** 03.009.25      **Issued on:** 30 January 2025  
**Last Amended on:** -      **Valid until:** 29 January 2028  
**Amendment no:** -

S.N.	Products/ Materials/ Items of test	Type of tests performed	Specifications/ Standard test methods/Techniques used	Range of testing/Limit of detection
<b>Field: Hematology</b>				
20.	Blood (Whole blood)	White Blood Cells (WBC)	Optical Flowcytometry	0–440×10 <sup>3</sup> /µL
21.	Blood (Whole blood)	Red Blood Cells (RBC)	Optical Impedance	0–8.6×10 <sup>6</sup> /µL
22.	Blood (Whole blood)	Hemoglobin	Photometry Cyanide-free SLS	0 –26 g/dL
23.	Blood (Whole blood)	Hematocrit (HCT)	RBC cumulative pulse height detection method	0.1–75%
24.	Blood (Whole blood)	Platelets (PLT)	Optical Impedance	0–5000×10 <sup>3</sup> /µL

\*\*\*END\*\*\*

  
 Quality Manager