BANGLADESH ACCREDITATION BOARD

ACCREDITATION CERTIFICATE

Issued under the authority of Bangladesh Accreditation Act, 2006
by Bangladesh Accreditation Board (BAB), Ministry of Industries to

Concord Diagnostics & Molecular Lab BD Ltd.

Navana Newbury Place, 4/1/A, (7th Floor)

Sobhanbag, Mirpur Road, Dhaka-1207, Bangladesh

This is to certify that this

Medical Laboratory

is accredited in accordance with the international standard

ISO 15189:2022

in respect of the associated scope, subject to the terms and conditions governing the relevant conformity assessment body (CAB) accreditation.

Certificate Number	:	**03.006.23**
Accreditation Date	:	**16 May 2023**
Date of Issuance	:	**04 December 2025**
Date of Expiration	:	**15 May 2026**

CERT. NO. 03.006.23
BAB
BANGLADESH ACCREDITATION BOARD
Accredited for ISO 15189:2022

GLOBAL ACCEPTANCE
BAB ACCREDITED
GLOBAL ACCEPTANCE

04.12.25

Mohd. Aminul Islam
Director General

This certificate must be returned on request; reproduction must follow BAB guidelines. For the specific scopes to which this accreditation applies, please refer to the Directory of CABs at BAB website.

BANGLADESH ACCREDITATION BOARD (BAB)

91, Motijheel C/A, Dhaka-1000
Tel: +880-2-9513221
Fax: +880-2-9513222
Email: info@bab.gov.bd
Web: www.bab.gov.bd

SCOPE OF ACCREDITATION

(For Testing Laboratory)

CAB Name & Address: **Concord Diagnostics & Molecular Lab BD Ltd.**
Navana Newbury Place, 4/1/A (7th Floor), Sobhanbag, Mirpur Road, Dhaka-1207, Bangladesh

Accreditation Standard: **ISO 15189:2022**
Accreditation Date: 16 May 2023
Certificate Number: 03.006.23
Issued on: 04 December 2025
Last Amended on: 04 December 2025
Valid until: 15 May 2026
Amendment no: 01

S. N.	Products/ Materials/ Items of test	Type of tests performed	Specifications/ Standard test methods/Techniques used	Range of testing/Limit of detection
Field: Clinical Chemistry				
1	Serum/Plasma	Albumin	Bromocresol Purple (BCP)	0.6 – 8.0 g/dL
2	Serum/Plasma	Total Protein	Biuret	2.0-12.0 g/dL
3	Serum/Plasma	Alkaline Phosphatase	PNPP, AMP Buffer-IFCC Ref.Proc.,Calibrated	10 – 1000 U/L
4	Serum/Plasma	SGPT/ALT	UV with P5P-IFCC Ref.Proc.,Calibrated	6 - 1000 U/L
5	Serum/Plasma	SGOT/AST	UV with P5P- IFCC Ref.Proc.,Calibrated	0 - 1000 U/L
6	Serum/Plasma	GGT	G-glutamyl-carboxy-nitroanilide	0-800 U/L
7	Serum/Plasma	Bilirubin (Total)	Diazonium Ion	0.1 - 25.0 mg/dL
8	Serum/Plasma	Bilirubin (Direct)	Diazotization	0.05-16 mg/dL
9	Serum/Plasma	Creatinine	Alkaline picrate, IFCC-IDMS standardized	0.15 - 20.0 mg/dL
10	Serum/Plasma	Urea	Urease, UV	0 - 321 mg/dL
11	Serum/Plasma	Uric Acid	Uricase, UV	0 - 20.0 mg/dL
12	Serum/Plasma	Glucose	Hexokinase	0 -27.8 mmol/L
13	Serum/Plasma	Total Cholesterol	Cholesterol oxidase, Esterase, Peroxidase.	50 - 600 mg/dL
14	Serum/Plasma	HDL – C	Direct measure-PEG	3 - 150 mg/dL
15	Serum/Plasma	LDL – C	Direct measure	5 - 300 mg/dL
16	Serum/Plasma	Triglyceride	Enzymatic, endpoint	15 -1000 mg/dL
17	Serum/Plasma	Na^+	ISE Indirect	50 – 200 mmol/L
18	Serum/Plasma	K^+	ISE Indirect	01 – 10 mmol/L
19	Serum/Plasma	Cl^-	ISE Indirect	50 – 200 mmol/L
20	Serum/Plasma	$HCO3^-$	Enzymatic	5 – 45 mmol/L
21	Serum/Plasma	Calcium	o-cresol phthalein complex one	5-15 mg/dL
Field: Immunology				
22	Serum	TSH	CLIA	0.01-150 µIU/mL
23	Serum/Plasma	FT4	CLIA	0.1 - 12ng/dL
24	Serum	FT3	CLIA	0.2 – 20 pg/mL
25	Serum	Total T4	CLIA	0.3 – 30 µg/dL
26	Serum	Prolactin	CLIA	0.3 – 200 ng/mL
27	Serum/Plasma	Testosterone	CLIA	7.0-1500 ng/dL

Quality Manager

BANGLADESH ACCREDITATION BOARD (BAB)

91, Motijheel C/A, Dhaka-1000
Tel: +880-2-9513221
Fax: +880-2-9513222
Email: info@bab.gov.bd
Web: www.bab.gov.bd

SCOPE OF ACCREDITATION

(For Testing Laboratory)

CAB Name & Address: **Concord Diagnostics & Molecular Lab BD Ltd.**
Navana Newbury Place, 4/1/A (7th Floor), Sobhanbag, Mirpur Road, Dhaka-1207, Bangladesh

Accreditation Standard: **ISO 15189:2022**
Accreditation Date: 16 May 2023
Certificate Number: 03.006.23
Issued on: 04 December 2025
Last Amended on: 04 December 2025
Valid until: 15 May 2026
Amendment no: 01

S. N.	Products/ Materials/ Items of test	Type of tests performed	Specifications/ Standard test methods/Techniques used	Range of testing/Limit of detection
28	Serum/Plasma	Ferritin	CLIA	0.5 – 1650 ng/mL
29	Serum	PSA	CLIA	0.01 – 100 ng/mL
30	Serum/Plasma	Vit. D	CLIA	4.2-150 ng/mL

END

Quality Manager