



ACCREDITATION CERTIFICATE

Issued under the authority of Bangladesh Accreditation Act, 2006
by Bangladesh Accreditation Board (BAB), Ministry of Industries to

Eurosia ITC Services Ltd.

Haveily Center (4th Floor), KA -11/2/A, Jagannathpur

Bashundhara Road, Vatara, Dhaka – 1229, Bangladesh

This is to certify that this

Calibration Laboratory

is accredited in accordance with the international standard

ISO/IEC 17025:2017

in respect of the associated scope, subject to the terms and
conditions governing the relevant conformity assessment
body (CAB) accreditation.

Certificate Number : 02.017.23
Accreditation Date : 23 May 2023
Date of Issuance : 16 June 2026 (1st Renewal)
Date of Expiration : 22 May 2029



A. Islam.
16.06.26
Mohd. Aminul Islam
Director General


This certificate must be returned on request; reproduction must follow BAB guidelines. For the specific scopes to which this accreditation applies, please refer to the Directory of CABs at BAB website.

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Eurosia ITC Services Ltd.
Haveily Center (4th Floor), KA -11/2/A, Jagannathpur, Bashundhara Road, Vatara, Dhaka – 1229.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 23 May 2023
Certificate Number: 02.017.23 **Issued on:** 16 June 2026
(1st Renewal)
Last Amended on: N/A **Valid until:** 22 May 2029
Amendment no: N/A

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
			50 g	0.27 mg
			100 g	0.48 mg
			200 g	1.09 mg
			500 g	0.09 g
			1 kg	0.09 g
			2 kg	0.10 g
			5 kg	0.12 g
			10 kg	0.14 g
			20 kg	0.16 g
04.	Micro Pipette	EISL-WI-04M	100 µl to 1000 µl	0.30 µl
			1000 µl to 5000 µl	0.65 µl
05.	Volumetric Glass Ware- (Pipette/ Burette/ Volumetric Flask/ Measuring Cylinder/ Beaker/ Pycnometer/ Conical Flask/Karl Fisher/ Auto Dispenser)	EISL-WI-01M	1ml to 10 ml	0.35 µl
			10 ml to 100 ml	0.65 µl
Field: Thermal				
06.	Temperature- (Thermometer / Temperature Sensor / Temperature Data Logger / Temperature Sensor / Transmitter / RTD / Thermocouple) with/without Indicator, Infrared	EISL-WI-05M	-20 to 300 °C	0.21 °C
			35 to 650 °C	0.25 °C



Quality Manager

